

# ARK™ Levetiracetam Control

This ARK Diagnostics, Inc. package insert for the ARK Levetiracetam Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

#### **Customer Service**



48089 Fremont Blvd Fremont, CA 94538 USA Tel: 1-877-869-2320 Fax: 1-510-270-6298

customersupport@ark-tdm.com

www.ark-tdm.com



Emergo Europe Westervoortsedijk 60 6827 AT Arnhem The Netherlands

CH REP MedEnvoy Switzerland

Gotthardstrasse 28 6302 Zug Switzerland

## Key to Symbols Used

LOT	Batch code	YYYY- MM-DD	Use by/Expiration date
REF	Catalog Number	<b></b>	Manufacturer
EC REP	Authorized Representative	CE	CE Mark
IVD	In Vitro Diagnostic Medical Device	<b>X</b>	Temperature limitation
Ţį.	Consult Instructions for Use	CONTROL	Quality Control
Rx Only	For Prescription Use Only		

### 1 Name

## **A**RK<sup>™</sup> Levetiracetam Control

### 2 Intended Use

ARK Levetiracetam Control is intended for use in quality control of the ARK Levetiracetam Assay, REF 5024-0001-00.

#### 3 Content

ARK Levetiracetam Control is comprised of a synthetic protein matrix with the following concentrations of levetiracetam:

REF	Product Description	Quantity/Volume
5024-0003-00	ARK Levetiracetam Control Levetiracetam, buffer, bovine serum albumin, and preservatives (nominal level)	Dropper vials
	LOW (7.5 μg/mL)	1 X 4 mL
	MID (30.0 μg/mL)	1 X 4 mL
	HIGH (75.0 μg/mL)	1 X 4 mL

Each laboratory should establish its own ranges for each new lot of controls.

## 4 Warnings and Precautions

- · For In Vitro Diagnostic Use.
- · Do not mix controls from different lot numbers.
- · Use each lot as a set.

#### 5 Instructions For Use

- · For a complete summary and explanation of the levetiracetam assay, refer to the package insert for the ARK Levetiracetam Assay.
- · Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

## 6 Limitations of Procedure

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

## 7 Trademarks

**ARK**<sup>™</sup> is a trademark of **ARK** Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.