








ARK™ SDMA Control

This ARK Diagnostics, Inc. package insert for the ARK SDMA Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

Customer Service


ARK Diagnostics, Inc.
 48089 Fremont Blvd
 Fremont, CA 94538 USA
 Tel: 1-877-869-2320
 Fax: 1-510-270-6298
 customersupport@ark-tdm.com
 www.ark-tdm.com

Key to Symbols Used

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Consult Instructions for Use		Temperature limitation
	Quality Control		

1 Name

ARK^L SDMA Control

2 Intended Use

ARK SDMA Control is intended for use in quality control of the ARK SDMA Assay.

3 Content

ARK SDMA Control is comprised of a synthetic protein matrix with the following concentrations of SDMA:

REF	Product Description	Quantity/Volume
5085-0003-00	ARK SDMA Control SDMA, buffer, bovine serum albumin, and preservatives (nominal level)	Dropper vials
	LOW (14.0 µg/dL)	2 X 2 mL
	HIGH (50.0 µg/dL)	2 X 2 mL

Each laboratory should establish its own ranges for each new lot of controls.

4 Warnings and Precautions

- For Veterinary Use only.
- Do not mix controls from different lot numbers.
- Use each lot as a set.
- Controls contain ≤0.09% sodium azide.

5 Instructions For Use

- For a complete summary and explanation of the ARK SDMA Assay, refer to the package insert for the ARK SDMA Assay, **REF** 5085-0001-00, **REF** 5085-0001-01 or **REF** 5085-0001-02.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date.

6 Limitations of Procedure

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 Trademarks

ARK^L is a trademark of **ARK** Diagnostics, Inc.
Other brand or product names are trademarks of their respective holders.



ARK Diagnostics, Inc.
Fremont, CA 94538 USA

Revised January 2026
1600-1214-00 Rev 02